

# Application of Multiobjective Genetic Algorithms in Anatomy Based Dose Optimization in Brachytherapy and its Comparison with Deterministic Algorithms

Natasa B. Milickovic<sup>1</sup>, Member, Michael Lahanas<sup>1</sup>, Maria Papagiannopoulou<sup>1</sup>, Kostas Karouzakis<sup>2</sup>, IEEE, Dimos Baltas<sup>1,2</sup>, Member, IEEE, Nikolaos Zamboglou<sup>1,2</sup>

<sup>1</sup>Dept. of Med. Physics & Eng., Klinikum Offenbach, Germany (e-mail: [Milickovic@aol.com](mailto:Milickovic@aol.com)),

<sup>2</sup>Institute of Communic. & Comp. Systems, National Technical University of Athens, Athens, Greece

## Abstract

In High Dose Rate (HDR) brachytherapy the conventional dose optimization algorithms consider the multiple objectives in form of an aggregate function which combines individual objectives into a single utility value. As a result, the optimization problem becomes single objective, prior to optimization. Up to 300 parameters must be optimized satisfying objectives which are often competing. We use multiobjective dose optimization methods where the objectives are expressed in terms of quantities derived from dose-volume histograms or in terms of statistical parameters of dose distributions from a small number of sampling points. For the last approach we compare the optimization results of evolutionary multiobjective algorithms with deterministic optimization methods. The deterministic algorithms are very efficient and produce the best results, but they also have the certain limitations. The performance of the multiobjective evolutionary algorithms is improved if a small part of the population is initialized by deterministic algorithms.

## 1. Introduction

High dose rate brachytherapy is a treatment method for cancer where empty catheters are inserted within the tumor volume. Once the correct position of these catheters is verified, a single <sup>192</sup>Ir source is moved inside the catheters at discrete positions (dwell positions) using a computer controlled machine. The problem that we consider is the determination of the  $n$  dwell times (which sometimes are called as well dwell position weights or simply weights) for which the source is at rest and delivers radiation at each of the  $n$  dwell positions, resulting in a three-dimensional dose distribution which fulfills the defined quality criteria. In modern brachytherapy, the dose distribution has to be evaluated with respect to the irradiated normal tissues and the Planning Target Volume (PTV) which includes besides the Gross Tumor Volume (GTV) an additional margin accounting for position inaccuracies, patient movements, etc. Additionally, for all critical structures, either located within the PTV or in its immediate vicinity or otherwise within the body contour, the dose should be smaller than a critical dose  $D_{crit}$ . In practice it is difficult, if not impossible to meet all these objectives. Usually, the above mentioned objectives are mathematically quantified separately, using different objective functions and then added together in various proportions to define the overall treatment objective function [1, 2].

The number of source positions varies from 20 to 300. It is therefore a high dimensional problem with competing objectives. The use of a single weighted sum leads to information loss and is not generally to be recommended, especially for non convex problems and for those cases where objectives have not the same dimensions and in addition maybe competing. An understanding of which objectives are competing or non-competing is valuable information. We therefore use multiobjective evolutionary algorithms in HDR brachytherapy. One algorithm is based on the optimization of dose-volume histograms (DVH), which describes the distribution of the dose within an object, or from these derived distributions. These distributions are evaluated for the PTV, the surrounding tissue and organs at risk from a set of up to 100000 sampling points [3]. The calculation of the DVH requires a considerable amount of time and for implants with 300 sources the optimization requires a few hours. Another limitation of this method is that a comparison with deterministic algorithms is not possible. We have therefore considered the optimization of

the dose distribution using as objectives the variance of the dose distribution on the PTV surface and within the PTV obtained from a set of 1500-4000 sampling points. These functions are convex and a unique global minimum exists.

In the past comparisons of the effectiveness of evolutionary algorithms have been made with either other evolutionary algorithms [4] or with manually optimized plans [1, 2]. We have compared the Pareto fronts obtained by multiobjective evolutionary algorithms with the Pareto fronts obtained by a weighted sum approach using deterministic optimization methods [5].

We use here only objectives where gradient based algorithms are superior. However, we must consider also critical structures partly inside the target or close to it which have to be protected by excessive radiation. Other objectives are the optimum position and the minimum number of sources. In such cases the gradient based algorithms can not be used.

## 2. Methods

### 2.1 Calculation of the Dose Rate

The dose rate around each of the small cylindrical shaped sources is dominated by the  $1/r^2$  term with modifications due to absorption and scattering in the surrounding material. The dose value  $d(r)$  at  $r=(x, y, z)$  is:

$$d(x) = \sum_{i=1}^{N_s} w_i K(r-r_i) \quad (1)$$

In (1)  $r$  is the position of the  $i$ th source and  $N_s$  the total number of sources.  $K(r-r_i)$  is the dosimetric kernel describing the dose rate per unit source strength at  $r$  from a source positioned at  $r_i$ . The dwell position weight  $w_i = S_k \cdot t_i$  is proportional to the strength  $S_k$  of the of the single stepping source, where  $t_i$  is the dwell time of the  $i$ th source dwell position [6]. Because of the high dose gradients a dose specification at a single point inside the PTV is not possible in interstitial brachytherapy. For this reason we use as a reference dose  $D_{ref}$  the average dose value at the PTV surface.

### 2.2 Dose-Volume Histogram Based Optimization Using the Conformal Index

A conformal Index (COIN) was proposed as a measure of implant quality and dose specification in brachytherapy [7]. This index takes into account patient anatomy, both of the tumor and normal tissues and organs.


We describe the dependence of the conformal index COIN on the choice of the reference dose value as the COIN distribution, see Fig. 1(b). Usually the dose values are

## Report Documentation Page

<b>Report Date</b> 25OCT2001	<b>Report Type</b> N/A	<b>Dates Covered (from... to)</b> -
<b>Title and Subtitle</b> Application of Multiobjective Genetic Algorithms in Anatomy Based Dose Optimization in Brachytherapy and its Comparison with Deterministic Algorithms		<b>Contract Number</b>
		<b>Grant Number</b>
		<b>Program Element Number</b>
<b>Author(s)</b>	<b>Project Number</b>	
	<b>Task Number</b>	
	<b>Work Unit Number</b>	
<b>Performing Organization Name(s) and Address(es)</b> Dept. of Med. Physics & Eng., Klinikum Offenbach, Germany		<b>Performing Organization Report Number</b>
<b>Sponsoring/Monitoring Agency Name(s) and Address(es)</b> US Army Research, Development and Standardization Group (UK) PSC 802 Box 15 FPO AE 09499-1500		<b>Sponsor/Monitor's Acronym(s)</b>
		<b>Sponsor/Monitor's Report Number(s)</b>
<b>Distribution/Availability Statement</b> Approved for public release, distribution unlimited		
<b>Supplementary Notes</b> Papers from the 23rd Annual International Conference of the IEEE Engineering in Medicine and Biology Society, October 25-28, 2001 held in Istanbul, Turkey. See also ADM001351 for entire conference on cd-rom.		
<b>Abstract</b>		
<b>Subject Terms</b>		
<b>Report Classification</b> unclassified	<b>Classification of this page</b> unclassified	
<b>Classification of Abstract</b> unclassified	<b>Limitation of Abstract</b> UU	
<b>Number of Pages</b> 4		

normalized to  $D_{ref}$  and are given either as fractions or percentages of  $D_{ref}$ .

The cumulative dose volume histograms of the PTV and the body for a rib implant is shown in Fig. 1(a). Due the rapid decrease of the dose-volume histogram (DVH) of the body a large number of sampling points is necessary in order to calculate with a high accuracy the DVH, the COIN distribution and the COIN integral at dose values close to the reference dose value and above. The COIN distribution from the DVHs of Fig. 1(a) and the COIN integral is shown in Fig. 1(b).



**Fig.1.** (a) Dose-volume histograms of the PTV and the body as a function of dose. (b) The corresponding COIN distribution. The shaded area to the right of  $D/D_{ref}=1.5$  is the COIN integral. The objectives are maximum COIN value at  $D=D_{ref}$  and minimum COIN integral for the avoidance of high dose values in the PTV and the surrounding tissue.

### 2.3 Dose Statistics Based Optimization

The DVH based optimization method requires a large number of sampling points for the computation of the histograms and the COIN distribution and therefore is computational expensive. We have developed a stratified sampling approach where the sampling points are non uniform distributed and which reduces the number of required sampling points by a factor of 5-10. Even then for implants with 200-300 sources the optimization time can reach 1-2 hours. A comparison of the performance with deterministic and gradient based algorithms is not practical or not even possible.

Therefore we consider another set of two objectives: For the conformity objective we use the variance  $f_s$  of the dose distribution of sampling points uniformly distributed on the PTV. In order to avoid excessive high dose values inside the PTV we require a small as possible dose distribution variance  $f_v$  inside the PTV. Due to the source characteristics these two objectives are competing. We use normalized variances for the two objectives:

$$f = \frac{1}{m^2 N} \sum_{i=1}^N (d_i - m)^2 \quad (4)$$

Where  $m$  is the average dose value and  $N$  the corresponding number of sampling points.

### 2.4 Multiobjective Optimization with Deterministic Algorithms

These objectives allow us to use deterministic gradient based algorithms. We use a weighted sum approach for the multiobjective optimization, where for a set of weights for the volume and surface variance we perform a single objective optimization of

$$f_w = w_s f_s + w_v f_v \quad (5)$$

where  $w_s, w_v \geq 0$  are the surface and volume importance factors, respectively and  $w_s + w_v = 1$ . We used 21

optimization runs where  $w_s$  varied from 0 to 1 in steps of 0.05 to determine the shape of the trade-off curve. A problem in using deterministic optimization methods is that the solution contains a large number of dwell weights with negative values. This is a non physical solution. In the past either constrained optimization methods were used or a correction was applied by setting to 0 all negative weights in each optimization step. A constrained optimization method increases the number of parameters by a factor of two. The correction method for the negative weights reduces the quality of the optimization results. We use a simple technique by replacing the decision variables, the weights  $w_k$ , with the parameters  $w'_k = w_k^{1/2}$ . Using this mapping technique we avoid non feasible solutions. For this unconstrained optimization we use the Polak-Ribiere variant of Fletcher-Reeves algorithm or the Broyden-Fletcher-Goldfarb-Shanno quasi-Newton based algorithm [5]. These require the first derivative of the objective function with respect to the decision variables to be calculated. The derivative of the normalized variance  $f$  used by the gradient based optimization methods is:

$$\frac{\partial f}{\partial w_k} = \frac{2}{Nm^3} \sum_{i=1}^N (d_i - m) \left[ m \frac{\partial d_i}{\partial w_k} - d_i \frac{\partial m}{\partial w_k} \right] \quad (6)$$

As a gradient free method we used the modified Powell method of Numerical Recipes [5].

### 2.5 Multiobjective Optimization with Evolutionary Algorithms

The population of multiobjective algorithms consists of strings storing a set of weights for each source dwell position. The weights are initially produced randomly distributed within the interval [0,1]. A part of population can be initialized by solutions of the deterministic algorithms, and more on this will be written further [8].

In our algorithm analysis we used these selection mechanisms:

- The niched Pareto algorithm (NPGA) proposed by Horn and Nafpliotis [9];
- Strength Evolutionary Approach algorithm (SPEA) by Zitzler and Thiele [10];
- Non-dominated Ranking Algorithm by Fonseca and Fleming (FFGA) [11, 12]
- Non-elitist Non-dominated Sorting Genetic Algorithm (NSGA) and Controlled Elitist Non-dominated Sorting Genetic Algorithms (NSGA II) [13,14,15]

After a new population is formed, the strings of randomly selected pairs undergo a crossover operation with a probability  $P_c$  and mutation with a probability  $P_m$ . We have found that  $P_c$  must be larger than 0.7 and  $P_m$  should be smaller than 0.1. The size of the population should be larger than 50. Various crossover types can be selected such as single point, two point, and arithmetic crossover. For the mutation operation also we have used various forms: uniform or non-uniform mutation. We use a real representation for the gene values. A detailed description of the genetic operators is given in reference [8].

### 2.6 Selecting the Solution from the Pareto Set

After the last generation is processed by the SPEA, FFGA, NPGA or NSGA II algorithm, members of the population are expected to be close to the Pareto frontier. A member of the non dominated set is selected which has a minimum Euclidean distance to the ideal optimum. The ideal point is defined by the minimum values  $(f_1^{min}, f_2^{min})$  of each objective function. The distance is calculated by normalizing each objective to a maximum value of 1 using the corresponding largest objective value found in the population. This member is presented as the

solution of the optimization process. Additionally members are selected each with the best result in each objective. A list is produced with the objective values for all the members of the Pareto set. Additionally the user can examine the dose distributions and the dose-volume histogram and isodose contours of every member of the population. Based on this information of the trade-off surface of the various objectives

a decision maker can select the best result.

### 3. Results

The dose variances are calculated from 1000 ÷ 4000 quasi-randomly distributed sampling points. For the COIN based optimization  $\approx 100000$  points are generated. The distances of these points to each source dwell position  $r$ , more precisely the inverse square distances  $1/r^2$ , are stored for speed maximization in look-up tables. We assume a invariant kernel  $K(r) = 1/r^2$  and ignore any spatial anisotropy, namely attenuation and scattering effect. This dosimetric simplification has no measurable influence on the results of the optimization. All calculations presented in our study have been made by using for the mutation probability  $P_m$  a value of 0.0065 and for the crossover probability  $P_c$  a value of 0.85.




Fig.2. Flow diagram for the DVH-based multiobjective genetic algorithm.




Fig.3. Comparison of the COIN distributions for a breast implant from the multiobjective genetic algorithm and four conventional single objective algorithms.

The flowchart for the COIN based optimization algorithm is shown in Fig. 2. For each member of the population for a given generation a renormalization is carried out according to the resulting COIN distribution, so that the maximum COIN value is observed at  $D = D_{ref}$  [7]. The dose prescription is realized at the  $D_{ref}$ , the isodose value resulting in the maximal conformity. This results generally in mean normalized dose values at the surface of PTV different from 1.0.

The multiobjective genetic algorithm, which uses dose-volume based constraints, produces equivalent or even better results than algorithms which were based on

phenomenological methods and used in the majority of treatment planning systems [16,17,18].

As an example in Fig. 3 the multiobjective genetic algorithm provides a solution with a more homogeneous dose distribution inside the PTV than by conventional optimization algorithms of a treatment planning system. Due to the large computational time for the COIN based optimization we used only the NPGA algorithm.

An example of the geometry of a PTV is shown in Fig. 5(a) Mic including the catheters, the source dwell positions and the sampling points on the PTV surface which define the surface variance. In Fig. 5(b) Mic the isosurface for the prescription dose is shown, which should have the same shape as the PTV.

The deterministic gradient based algorithms are very effective in generating the Pareto front using a summed weights approach. Powells algorithm which does not require derivatives is efficient only for implants with a small number of sources (time consuming), whereas the gradient based algorithms require only 1-2 minutes. Gradient based algorithms are limited by the fact that they

can be trapped in local minima, or that non convex regions are not accessible using the weighted sum method [19].

From the evolutionary algorithms NSGAI and SPEA have been found to produce the best results, since it applies an elitism and sharing mechanism. For implants with many sources the genetic algorithms used converge in some cases to a Pareto set which was far away from the true Pareto set. Such an example for an implant with 215 source dwell positions is shown in Fig. 5. The SPEA algorithm converges after 200 generations to a Pareto front which is very small and far from the Pareto set generated by the gradient based algorithms. The optimization path is shown for a set of importance factors  $f_v$ ,  $f_s$  for the Polak-Ribiere algorithm. After 10 iterations a point on the Pareto front is reached.

Using random sets of decision variables we have found for this example that the number of function evaluations required by a random search method to obtain points on the Pareto front




Fig.4. (a) Contours of a rib implant with the catheters and the source dwell positions. On the PTV surface sampling points are shown at which the dose is calculated (b) the dose isosurface obtained from the dose optimization.




Fig.5. Pareto front obtained by the gradient based algorithm and with the SPEA algorithm with and without initialization.

is larger than  $10^{30}$  [8]. A random search would require  $10^{10}$  times more function evaluations to generate points on the Pareto set found by the SPEA algorithm without initialization. Even with this performance the SPEA algorithm is not able to produce points on the Pareto front found by the deterministic methods. Using a few members initialized by the gradient based algorithm the multiobjective evolutionary algorithms reproduced the Pareto fronts obtained by the deterministic algorithms, Fig. 5. For a more detailed comparison of the deterministic and evolutionary algorithms see reference [8].

#### 4. Conclusions

We used for the first time multiobjective evolutionary anatomy based dose optimization algorithms in HDR brachytherapy [6]. For the COIN-based objectives we have found that multiobjective evolutionary algorithms produced solutions which are better than by conventional algorithms in treatment planning systems which use deterministic algorithms and catheter-oriented objectives. They also have the problem with infeasible negative weights which they avoid by a repair mechanism or by using special constraints to the objective functions in order to reduce their numbers and the degree of the violation.

The results of various algorithms for the variance based objectives have been compared using a representative set of 22 implants encountered in clinical practice. We have limited our study to cases where no critical structures are considered. Trade-off surfaces which reveal the nature of the multiobjective problem of the dose optimization in brachytherapy have been obtained. Due to the variety of the trade-off surfaces found, which depends on the implant and complex catheter geometry, no common set of optimal importance factors exists. Therefore it is useful to determine the Pareto front and then to select a solution according to its characteristics. Pareto sets have been obtained by a deterministic unconstrained optimization method using a simple mapping technique which transforms the linear into a quadratic optimization problem and removes infeasible solutions with negative dwell position weights. The gradient based algorithms, if they can be used, are very effective because they converge very fast and generate the Pareto fronts which in most cases are much better than the Pareto front obtained by evolutionary multiobjective algorithms.

If the number of objectives increases then the number of combinations using a weighted sum approach with deterministic algorithms increases. Deterministic methods are not efficient for non analytic complex objectives such as used by the COIN based method. When more objectives are included then a non convex feasible space could be the result [20]. A combination of deterministic and evolutionary multiobjective algorithms seems to be the best choice for a robust and efficient multiobjective dose optimization in HDR brachytherapy.

We are currently studying for various sets of objectives the Pareto fronts using multiobjective evolutionary algorithms and if possible in combination with deterministic algorithms.

#### References

1. Yu, Y., Schell, M. C., "A genetic algorithm for the optimization of prostate implants", *Med. Phys.* 23, (1996) 2085-2091
2. Yang, G., Reinstein, L. E., Pai, S., Xu, Z., "A new genetic algorithm technique in optimization of permanent  $^{125}\text{I}$  prostate implants", *Med. Phys.* 25 (1998) 2308-2315
3. Lahanas, M., Baltas, D., Giannouli, S., Milickovic, N., Zamboglou, N., "Generation of uniformly distributed dose points for anatomy-based three-dimensional dose optimization methods in brachytherapy", *Med. Phys.* 27 (2000) 1034-1046
4. Zitzler, E., Deb, K., Thiele, L., "Comparison of Multiobjective Evolutionary Algorithms", *Empirical Results. Evolutionary Computation*. 8 (2000) 173-195
5. Press, W. H., Teukolsky, S. A., Vetterling, W. T., Flannery, B. P., *Numerical Recipes in C*. 2nd ed. Cambridge University Press, Cambridge, England. 1992
6. Nath, R., Anderson, L. L., Luxton, G., Weaver, K. A., Williamson, J. F., Meigooni, A. S., "Dosimetry of interstitial brachytherapy sources", Recommendations of the AAPM Radiation Therapy Committee Task Group No. 43. *Med. Phys.* 22 (1995) 209-234
7. Baltas D., Kolotas, C., Geramani, K., Mould, R. F., Ioannidis, G., Kekchidi, M., Zamboglou, N., "A Conformal Index (COIN) to evaluate implant quality and dose specifications in brachytherapy", *Int. J. Radiat. Oncol. Biol. Phys.*, 40 (1998) 512-524
8. Milickovic, N., Lahanas, M., Baltas, D., Zamboglou, N., "Comparison of evolutionary and deterministic multiobjective algorithms for dose optimization in brachytherapy", *Evolutionary Multi-Criterion Optimization, Lecture Notes in Computer Science* 1993, Zitzler, E., Deb, K., Thiele, L., Coello Coello, C. A., Corne, D. Eds., EMO 2001, Zurich, Switzerland, March, 2001
9. Horn, J., Nafpliotis, N., "Multiobjective optimization using the niched Pareto genetic Algorithm", *IlligAL Report No. 93005*. Illinois, Genetic Algorithms Laboratory, University of Illinois at Urbana-Champaign, 1993
10. Zitzler, E., Thiele, L., "Multiobjective Evolutionary Algorithms: A Comparative Case Study and the Strength Pareto Approach", *IEEE Transactions on Evolutionary Computation*. 37 (1999) 257-271
11. Fonseca, M., Fleming, P. J., "Multiobjective optimization and multiple constraint handling with evolutionary algorithms I: A unified formulation", *Research report 564*, Dept. Automatic Control and Systems Eng. University of Sheffield, Sheffield, U.K., Jan. 1995
12. Fonseca, M., Fleming, P. J., "An overview of evolutionary algorithms in multiobjective optimization", *Evolutionary Computation* 3 (1995) 1-16
13. Ref. Srinivas N and Deb K, "Multi-objective function optimization using non-dominated sorting genetic algorithms", *Evolutionary Computation* 2 (1995) 221-248.
14. Deb K, Pratap A, Agrawal S., Meyarivan T, "A fast and elitist multiobjective genetic algorithm: NSGA II", *Technical Report No. 2000001*, Kanpur: Indian Institute of Technology Kanpur, India, (2000).
15. Deb K, Agrawal S., Pratap A, Meyarivan T, "A fast and elitist non-dominated sorting genetic algorithm for multiobjective optimisation: NSGA II", *Proceedings of the Parallel Problem Solving from Nature VI Conference*, pp. 849-858. (2000)
16. Lahanas, M., Baltas, D., Zamboglou, N., "Anatomy-based three-dimensional dose optimization in brachytherapy using multiobjective genetic algorithms", *Med. Phys.* 26 (1999) 1904-1918
17. Edmundson, K., "Geometry based optimization for stepping source implants, in: *Brachytherapy HDR and LDR*", A. A. Martinez, C. G. Orton and R. F. Mould eds., *Nucletron: Columbia*. (1990) 184-192
18. Van der Laarse, T. P. E. Prins., "Introduction to HDR brachytherapy optimization", In: R. F. Mould, J. J. Battermann, A. A. Martinez and B. L. Speiser eds. *Brachytherapy from Radium to Optimization*. Veenendaal, The Netherlands: Nucletron International. (1994) 331-351
19. Das, I. Dennis, J., "A Closer Look at Drawbacks of Minimizing Weighted Sums of Objectives for Pareto Set Generation in Multicriteria Optimization Problems", *Structural Optimization* 14 (1997) 63-69
20. Deasy, J. O., "Multiple local minima in radiotherapy optimization problems with dose-volume constraints", *Med. Phys.* 24 (1997) 1157-1161